

Getting Started

Default Recorder Settings

- 7-day chart rotation
- +32 to 120°F (TH800/02)
- 20 to +120°F (TH803/05)

Quick Start

1. Remove the protective pen caps.
2. Insert batteries and plug in AC adapter.
(See Figure 3.) (TH8 requires 4 AA batteries only.)
The unit will power on.
3. The instrument will move pen(s) to the appropriate readings.
4. Install the chart that matches the Dip Switch settings.
Press the **Pen Home** key to move the pens to the outside of the chart. The pen is automatically raised off the chart.

Remove the old chart, place the new chart on the *Chart Hub* being certain that the edge of the chart slides under the *Chart Guide Clips* located at the outside of the chart.

5. **Set the appropriate time.** There are two ways to adjust the chart and set the appropriate time:
 - a. Set the chart time manually by inserting a coin into the groove in the chart hub and turning clockwise until the correct hour (and day if applicable) on the chart is referenced to the timing arrow (just to the right of the pen tip on the dial).

This feature should be used for fine adjustments only.

- b. To adjust the chart time, press and hold **Adjust-Up** and **Adjust-Down** buttons located on the back of the unit next to the dip switch (see Figure 3).

The green LED will blink rapidly for about five seconds, then the LED will remain solid green. While in this state the **Adjust-Up** button will move the chart backward (counter clockwise) and the **Adjust-Down** button will move the chart forward (clockwise).

Rotate the chart until the correct hour (and day if applicable) on the chart is referenced to the timing arrow. Once you have set your chart, press the **Pen Home** button to exit *Chart Adjust Mode*. The unit will take one minute to exit the *Chart Adjust Mode* once **Pen Home** is pressed.


Press **Pen Home** to move the pens back onto the chart.

6. Place the TH8 recorder on a flat vibration-free surface. Be sure it is in a vertical position and level. For best performance and longevity, the location should be a clean environment, free from dust and corrosive fumes. Do not exceed temperature specifications.

Wall Mount Keyhole slots are provided on the TH8 recorder for wall mounting.

Display Symbols TH8

Figure 1




For setting alarms only


Button Operations

Figure 2

TH802/03/05



TH800



On/Off

The On/Off key turns the unit on and off.

Pen Home (Down Arrow)

If the pens are located on the outside edge of the chart, press the **Pen Home** to move pen(s) to recording position. If the pens are located on the chart, press the **Pen Home** key to move the pen(s) to the outside edge of the chart.

Alarm Optional

Pen 1: = Red Pen

Pen 2: = Blue Pen

1. To set the alarm, make sure unit is on and press and hold the **Alarm** button. The LED will blink red rapidly for about five seconds, then the LED will turn solid green. Release the **Alarm** button and the LED will turn solid red. At this point the display will show "On" or "Off".

Pressing either the **Adjust-Up** or **Adjust-Down** buttons, located on the back of the unit next to the dip switch, will toggle the alarm on or off.

2. Pressing **PEN HOME** will scroll to the next alarm options.

Alarm "on" or "off"




Figure 3

- Pen 1 alarm minimum
- Pen 1 alarm maximum
- Pen 2 alarm minimum
- Pen 2 alarm maximum

3. In order to set pen alarm minimums and maximums, pressing **Adjust-Up** will increase the alarm value, and pressing **Adjust-Down** will decrease the alarm value. There is acceleration if the **Adjust-Up** button is held down.

Repetitively pressing the **Pen Home** button will scroll through the five options until the alarm button is pressed to exit from alarm adjust. Each press of the **Pen Home** or **Alarm** button will store the new settings. The unit will take one minute to exit **Alarm Set Mode** once the **Alarm** button is pressed.

4. If the alarm is triggered, the LED will show as solid red and the alarm will sound. Press the **Alarm** button to silence the audible alarm.


Alarm Relays

The SPST 24V 500mA relay contacts are normally open and will close on alarm conditions when the alarms are enabled. Relay contacts are always functional when the alarm is enabled. The relay will close only during minimum and maximum alarm conditions. Two pen models have 2 sets of relay contacts which operate independently. Relays will not operate under battery only power conditions.

Dip Switch Setup

To set up the **TH8** recorder for your specific application, you might need to change some of the "Dip Switches". The Dip Switches are located on the back of the unit. (Figure 3)

Use a pen or small screw driver to flip the switches. *Remember to install the correct chart to match the corresponding Dip Switch setting.*



Dip Switches for Ranges

TH8 Dip Switch Setting		Up=On Down=Off					
Switches 1-6	For Model	1	2	3	4	5	6
7-Day	All Models	D	D				
24-Hour	All Models	D	U				
31-Day	TH802/03/05	U	U				
RH	All Models						D
Dew Point	All Models						U
-20° to 120°F	All Models			D	D	D	
40° to 110°F	TH802/03/05			D	D	U	
0° to 200°F	TH803/05			D	U	U	
-20° to 50°C	All Models			U	D	D	
5° to 40°C	TH802/03/05			U	U	D	

Remember to install the correct chart to match corresponding switch setting

DICKSON • Addison, IL • www.DicksonData.com

Note: TH800 -On Dip Switch label -20 to 120°F = +32 to 120°F and -20 to 50°F = 0 to 50°C

Recording Time

All the TH8 recorders feature user selectable recording times.

24-hour all models	#1 down
	#2 up
7-day all models	#1 down
	#2 down
31-day (TH802/03/05)	#1 up
	#2 up

Temperature Ranges

-20 to +120°F	#3 down
	#4 down
	#5 down
+40 to +110°F	#3 down
	#4 down
	#5 up
-20 to +50°C	#3 up
	#4 down
	#5 down
+5 to +40°C	#3 up
	#4 up
	#5 down
0 - 200°F	#3 down
	#4 up
	#5 up

RH/Dew Point

Dip switch #6 allows you to select whether RH or dew point is recorded by the blue pen and shown on the display for the models with displays. (The red pen always records temperature.)

RH	#6 down
Dew Point	#6 up

Note: When the blue pen is recording dew point, the reading indicates the temperature at which dew will form or frost point, so you should use the temperature scale on the chart to read dew point.

Power

The **TH8** operates on AC power with an optional 4 AA battery backup. Depending on chart rotation selected and model; battery backup will last

TH802/03/05

24-Hour Chart Rotation = 1 Day Backup

7-Day Chart Rotation = 2 Day Backup

31-Day Chart Rotation = 1 Week Backup

Note: Alarm and relays will not operate under battery only power.

TH800

At 70°F

24-Hour Chart Rotation = 1 Month

7-Day Chart Rotation = 6 Months

At 32°F

24-Hour Chart Rotation = 3 Weeks

7-Day Chart Rotation = 4 Months

Note: (Unit should not be placed in an environment below 32°F)

LED Indicators

- AC Power with battery backup—Solid green
- AC Power with low battery or no battery—Blinks red
- Battery only—Blinks green
- Battery only (Low battery)—Solid red

Note: **TH800** operates on battery only and will blink one time every minute to conserve battery life.

CALIBRATION

Your instrument was carefully tested and calibrated before being shipped from the factory. For greatest accuracy, we recommend factory re-calibration every 6–12 months. Call customer service at (630) 543-3747.

If you wish to do calibration yourself, follow these procedures.

1. To activate Calibration Mode, turn the unit on and press and hold both the **On/Off** button and the **Adjust-Down** button (making sure to press the **Adjust-Down** button first so the unit will not turn off). The LED will blink in amber rapidly for about five seconds, and turn solid green. Release the **On/Off** button and the **Adjust-Down** button. The

LED will then blink amber at which point only the pen being adjusted will show on the display.

2. To raise the unit of measurement, press the **Adjust-Down** button. To lower the unit of measurement, press the **Adjust-Up** button. **Pen home** switches between the red and blue pens on a two-pen unit and stores the current adjustment value.
3. When calibration is complete, press the **On/Off** button. Calibration is stored in memory even after you turn the unit off. User calibration information will not be lost if AC power fails.

Note: After two hours, if no buttons are pressed, the unit will time out of user calibration and resume normal operation. The user calibration is stored separately from the factory calibration. If you wish to cancel out your user calibration, simply enter *Calibration Mode* and toggle through the steps without adjusting displayed readings. Exit by pressing the **On/Off** button. You have now restored factory calibration.

Note: It is recommended that you use a controlled chamber when calibrating temperature and humidity. Calibrating in an open room is not recommended as humidity can vary greatly within a very small area.

NEED TO KNOW

TH800 Pen Movement

The pen movement for the **TH800** is slower than on the AC powered models in order to conserve battery power. In 24-hour mode the pen will move every 1 minute and 30 seconds. In 7-day mode the pen will move every 8 minutes and 30 seconds.

The recorder takes temperature readings every 50 seconds. An average of the readings taken between pen movements is what is drawn on the chart.

Pen Adjust

Pen adjust may be needed if the pen(s) and display do not match.

1. While the unit is on, press and hold both the **Pen Home** button and the **On/Off** button (making sure to press the **Pen Home** button first so the unit will not turn off). The led will blink rapidly between amber and green for about five seconds. The LED will then turn green. Release the **Pen Home** and **On/Off** buttons. The LED will flash amber and green for one second and the LED will turn off.

2. Both pens will move to the outer edge of the chart, and then the shorter red pen will move to the current position. Use the **Adjust-Up** and **Adjust-Down** buttons to move the pen to the maximum temperature line on the chart. Once the shorter red pen is set, the **Pen Home** button will move the shorter red pen to the outer edge of the chart and the longer blue pen will move to its current position. The **Adjust-Up** and **Adjust-Down** buttons will adjust the longer blue pen position. Once the pen is set, the **Pen Home** button will move the longer blue pen to the outer edge of the chart and the shorter red pen will move to its new position. Each time pen home is pressed the two pens will switch positions (moving the most recently adjusted position) until the **On/Off** button is pressed to exit pen adjust mode.

Note: Pressing the **On/Off** button will exit the adjustment and return the unit to normal operation. Each time the **Pen Home** or **On/Off** button is pressed, the adjustment currently in progress will be stored. The unit will take one minute to exit *Pen Adjust Mode* once the **On/Off** button is pressed.

TROUBLESHOOTING

Why isn't the chart keeping time or running slow?

- The chart may be "hung up" or restricted, possibly caused by a rip on the outer edge of the chart or the chart hub, or the chart may be caught between the arm and pen arm platform.
- Incorrect chart installed for the selected chart speed.

Why did the chart stop turning?

- Chart hung up or restricted, (ripped chart)
- Unit may be "locked up"? This can be confirmed by pressing any of the buttons on the keypad. If the unit is locked up there will be no response to button presses and the Chart Recorder may appear to be working, but the readings won't change. Also, the chart will not rotate. Remove power and battery, and then re-power.

Why don't the display and chart match?

- Dip switches are set for a specific range, but using chart for another range, or vice versa.
- Pen(s) not inserted on pen arm(s) all the way
- To adjust pen to match chart see **Pen Adjust** in the *Need to Know* section of this manual. **Note:** TH800 does not have a display.

Why does calibration seem to be off?

- What is the tolerance of the unit it's being compared to?
- It's ok if the unit is within the sum of the two tolerances.
- Has an outside calibration house attempted calibration? It may not have been adjusted properly.
- Walk through calibration adjustment, found in the manual

Why won't the battery backup work?

- Are there good batteries in the Chart Recorder?
 - Keep in mind that battery backup varies significantly depending on temperature, pen movement and chart rotation speed.
 - If there was a momentary power outage (brown out), the unit may not have had enough time to recognize this and switch to *Battery Mode*. (The Chart Recorder may lock up or shut off in this situation) This situation may also occur if the unit is plugged into an outlet that shares the circuit with other machinery that has phase motors or compressors that cycle periodically. As these other machines cycle, they momentarily have a high current draw, therefore drawing from the unit.
- Note:** TH800 does not have a battery backup.

Why won't the chart recorder respond to changes?

- Is the unit "locked up"? This can be confirmed by pressing any of the buttons on the keypad. If the unit is locked up there will be no response to button presses and the unit may appear to be working, but the readings won't change. Also, the chart will not rotate. Remove power and battery, and then re-power.

Why won't the unit power up?

- Remove the batteries and power adapter for a minute or two, this will reset the unit. The unit should respond when the adapter is plugged in.

Warranty

Dickson warrants that this line of instruments will be free from defects in material and workmanship under normal use and service for a period of twelve months after delivery.

This warranty does not cover routine calibration and battery replacement.

For *Specifications* and *Technical Support* go to www.DicksonData.com

Factory Service & Returns

Contact *Customer Service* **630.543.3747** for a *Return Authorization Number* (RA) before returning any instrument. Please have the model number, serial number and a PO ready before calling.

www.DicksonData.com

930 South Westwood Avenue • Addison, IL 60101-4917

Telephone **630.543.3747** • Fax **630.543.0498** • Email **DicksonCSR@DicksonData.com**

Rev. 08/09